

CLARIFICATIONS

Contract title: public contract for the supply, delivery and installation of skills lab equipment at 3 health-training institutions as well as training of end users.

Reference: UGA22010-10007

Please note that the contracting authority will not further respond to any other question as provided for under point 3.3 "Information". As a reminder, tenders must be received **before 09th July 2024 at 12:00 Kampala time. Late bids will not be accepted.**

| N° | Issue / question raised | Response |
|----|---|--|
| 1. | Please advise Volume in Liters for the autoclaves | 20 Litres |
| 2. | Item 21 Instrument set, Anaesthesia QTY x10 is the same as Item 31 Instrument set, Anaesthesia Qty x10 Is this Intended or a mistake? | It is a repetition, quote for Item 21 Instrument set, Anaesthesia QTY x10 and leave out item 31 Instrument set, Anaesthesia. |