

CLARIFICATIONS

Contract title: Public contract for the Supply & delivery of Medical Spare Parts for Fort Portal

Regional Referral Hospital & Jinja Regional Referral Hospital

Reference: UGA22009-10013

Please note that the contracting authority will not further respond to any other question as provided for under point 3.3 "Information". As a reminder, tenders must be received before **23rd August 2024 at 2:00 PM Kampala time. Late bids will not be accepted.**

N°	Issue/question raised	Response		
1.	What is the model name or number for Autoclave 100L under LOT 1?	The model is Ls-B100L; Note its control mechanical		
2.	What are the part numbers of the Datex Ohmeda and UAM Gradian anaesthesia under LOT 2?	Description of equipment	Description spares needed	Part Numbers
		Anaesthetic machine Datex Ormeda model; 9100C [15 pcs]	Breathing system interface seals	1470-3320-000
			Vaporizer manifold seals 14.3mm ID	1102-3016-000
			Seal Ventilator Below	1006-3690-000
			Free-breathing value Flapper	0211-1454-100
			Free-breathing Valve O ring 34.59ID X 39.83OD	1503-3208-000
			FREE breathing valve O ring, 30IDX33.6 OD	M1168794-S
Flow Sensor assembly	M1174442-S1			

			Flow Sensor elbow connectors	M1171751-S
			Bellow	1500-3378-000
			Exhalation Diagram assembly	1503-8121-000
			Oxygen sensor including O ring	M-10
			Cable Oxygen SENSOR Cell	1009-5570-000
			Battery 12V 7AH Gel	M1168045-S
			Vaporizer Isoflurane	N/A
			Breathing Circuit Adult	N/A
			Breathing Circuit Paediatric	N/A
			Castors for GE anaesthesia Machine	N/A
			Soda Lime	N/A
			Breathing Bag Adult	N/A
			Breathing Bag Paediatric	N/A
			Bacteria filter	N/A
			Breathing Masks Adult	N/A
			Breathing mask Paediatric	N/A
			T-piece's	N/A
			High-pressure regulator	N/A
			Spindle Keys	N/A

			Oxygen Horse Pipe	N/A
		Universal Anesthetic Machine [UAM] Gradian Health System	Sieve beds	Concentrator sieve beds
			Service kits	Concentrator compressor service kit
			Transformer 12V 13A	Concentrator transformer
			Breathing circuit	UAM breathing circuit
			Flow Sensor	UAMV flow sensor
			Oxygen Sensor	Oxygen analyser sensor
			Oxygen Hose pipe	N/A
			Air filter	Concentrator air filter
			Castors for UAM	UAM caster [Front and Rear]
			Oxygen masks adult	N/A
			Oxygen masks Paediatric	N/A
			Breathing bags Adult	N/A
			Breathing bags Paediatric	N/A
			Battery 6V, 1A	Oxygen analyzer battery
			Scavenging tube	N/A
		Isoflurane Vaporiser	UAM Vaporiser Isoflurane	
3.	Is the oxygen delivery system 2L and 5L flowmeter for oxygen concentrators or wall outlets/cylinder outlets?	The delivery system is Oxygen concentrators		