

CLARIFICATION RESPONSES

Contract title: Public contract for the supply, delivery, installation and testing of medical equipment for health facilities as well as conducting user training.

Reference: UGA22009-10015

N°	Issue/question raised	Response
1.	Under LOT 1: Item No. 2 - Vacuum Extractor; vacuum extractor cup size: Ø50mm, 60mm & 940mm, It seems to be typing error. Generally, 940mm cup, which are available in size Ø40mm, 50mm, 60mm or, Ø50mm, 60mm, 70mm. Please advise.	Kindly consider the sizes of Ø50mm, 60mm, and 70mm.
2.	Under LOT 1: Item No. 5 - Universal Anaesthesia Machine (UAM) – The required anaesthesia machine is dedicated to one manufacturer. The different manufacturers have slightly different technical parameters but ultimately meet the desired output equivalent. What are the general specifications?	What is very important to this unit is that the machine should generate its oxygen. No option of external connection to the machine as a stand-alone will be accepted. Note An offer higher than the specification and that meets the desired output will be accepted.
3	Under LOT 1: Item No. 7 - Operation Table, Hydraulic /theatre bed – Universal operation table, mobile with simple mechanical and hydraulic system, Height adjustment (mm): 650 to 1,040 & Trendelenburg 45 & Anti- Trendelenburg 20, the required height adjustment and Trendelenburg angle is not possible collectively on mechanical and hydraulic system OT table, even on electromechanical / electric table can not reach 45° Trendelenburg. Please advise.	This is noted. Kindly consider the minimum value of 20° for the Trendelenburg & Anti- Trendelenburg angle

4.	<p>Under LOT 1:</p> <p>Item No. 9 - Delivery Bed - As per specification required delivery bed leg section completely recessed in body section. It is also written that leg section should partially be extendable to serve as a shelf for receiving infant or basin which seems to be a mismatch with the standard bed. Please clarify.</p>	<p>This facilitates movement of the leg section underneath the head section. Note the leg section is on castors.</p>
5.	<p>Under LOT 2:</p> <p>Item No. 4 - Phototherapy Unit - Standards and Certifications - CE Mark, FDA (US) /STQC CB certified, STQCS certified or supplied with a functional safety test report from ERTL.</p> <p>The requested certification seems to be of a specific manufacturer. Could you please generalise it to ISO, CE.</p>	<p>This is acceptable ISO, and CE marks.</p>
6.	<p>Under LOT 2:</p> <p>Item No. 7 – CPAP – Reference is made to clarification no. 2 – point 8 about CPAP. The brand available in the local market will not give Manufacturer authorization to other bidders. Please provide general specifications of CPAP to allow alternate competitive options.</p>	<p>The general specifications will remain. However, for an integrated pulse oximeter, an option for a handheld pulse oximeter, or integrated pulse oximeter will be accepted. For which ever option the tenderer chooses to provide, a MAF will be required</p>